



## **Introduction to Particular Standards of Plasma Medical Equipment for Wound Healing**

**Jinsung Choi<sup>1</sup>, Junsup Lim<sup>2</sup>, Sehun Ki<sup>2</sup>, Young June Hong<sup>1</sup>, Eun Ha Choi<sup>1,2</sup>**

<sup>1</sup>Plasma Bioscience Research Center, Kwangwoon University, 447-1, Seoul 01897, Korea

<sup>2</sup>Department of Electronic and Biological Physics, Kwangwoon University, 447-1, Seoul 01897, Korea.

E-mail: [neoled@kw.ac.kr](mailto:neoled@kw.ac.kr)

Standard is an important in the development of devices using plasma. There is no national or international standard for plasma equipment. It introduces the particular standard of Plasma medical equipment for wound healing. This standard is prepared by referring to the guidelines of the Korea Food and Drug Administration and the DIC SPEC 91315 of Germany. The suggested test methods have been carried out with the plasma devices of PBRC and Korean companies. This standard includes a method for Confirmation of plasma and its species for wound healing plasma medica equipment, protection against excessive reactive species, protection against electrical HAZARD and protection against excessive temperatures for Plasma gas.